

Document reference no.: 152515/1

Report no.: RN-0290





EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

HI98195

Multiparameter pH/ ORP/ EC/ Pressure/ Temperature Waterproof Meter

Are in compliance with:

Directive 2011/65/EU

Directive 2014/30/EU
Directive 2014/35/EU

RoHS2

EMC and its standards Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or reference s to the other technical specification:

EN 50581:2012

Restriction of the use of certain hazardous substances (RoHS)

EC Regulation 1907/2006

Registration, Evaluation, Authorization and Restriction of Chemicals (REACH) Electrical Equipment of Measurement, Control and Laboratory use

EN 61326-1: 2013

- EMC requirements Part 1: General requirements and its standards:

EMC requirements Part 1:

General requirements and its standards:

EN 61000-4-2: 2009, EMC EN 61000-4-3: 2006, EMC Electrostatic Discharge Immunity Test Radiated RF, Electromagnetic Field Immunity Test 4 kV / 8 kV contact / air 3 V / m (80 Mhz to 1 GHz)

Class A

EN 61010-1: 2010 (IEC 61010-1: 2010)

Safety Requirements for Electrical Equipment for Measurement, Control and

Laboratory Use Part 1: General Requirements.

 $The \ manufacturer \ is \ not \ liable \ for \ modifications \ made \ by \ the \ user \ which \ may \ alter \ the \ device's \ conformity \ with \ the \ CE \ label, \ nor \ the \ consequences \ thereof.$

This declaration is the responsibility of the manufacturer:

S.C. Hanna Instruments S.R.L.

Str. Hanna Nr.1, loc. Nușfalău, jud. Sălaj, cod 457260,

ROMANIA

Hanna Instruments, Inc.

Highland Industrial Park, 584 Park East Drive,

Woonsocket, RI 02895, USA Phone: +1 401 7657500



Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Hanna Instruments SRL

Test performed by:

Jozsa Ervin / ETO Eng. /

[Name / Title of Signatory / Signature]

Declaration authorized by: Vasile Prodan / ETO Responsible /

[Name / Title of Signatory / Signature]

Issue place / Date

Nuşfalău / February 07, 2018

[Place / Date in format: Month, dd, yyyy]

FO-CETEST r.3 - 2018,01,30

Page 1 Of 1