

Document reference no.: 154612/3

Report no.: RN-0332b



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

HI720	Calcium Hardness Colorimeter
HI721	Iron Colorimeter
HI723	High Range Chromium VI Colorimeter
HI736	Saltwater Aquarium Ultra Low Range Phosphorus Colorimeter
HI749	Low Range Chromium VI Colorimeter
HI761	Ultra Low Range Total Chlorine Colorimeter
HI762	Ultra Low Range Free Chlorine Colorimeter
HI764	Saltwater Aquarium Ultra Low Range Nitrite Colorimeter
HI771	Ultra High Range Total Chlorine Colorimeter

Are in compliance with:

Directive 2011/65/EU	RoHS2
Directive 2014/30/EU	EMC and its standards
Directive 2014/35/EU	Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or reference s to the other technical specification:

EN 50581:2012	Restriction of the use of certain hazardous substances (RoHS)	
EC Regulation 1907/2006	Registration, Evaluation, Authorization and Restriction of Chemicals (REACH)	
EN 61326-1: 2013	Electrical Equipment of Measurement, Control and Laboratory use	Class A
	- EMC requirements Part 1: General requirements and its standards:	
EMC requirements Part 1:	General requirements and its standards:	
EN 61000-4-2: 2009, EMC	Electrostatic Discharge Immunity Test	4 kV / 8 kV contact / air
EN 61000-4-3: 2006, EMC	Radiated RF, Electromagnetic Field Immunity Test	10 V / m (80 Mhz to 1 GHz)
EN 61010-1: 2010 (IEC 61010-1: 2010)	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.	

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

S.C. Hanna Instruments S.R.L.

Str. Hanna Nr.1, loc. Nușfalău,
jud. Sălaj, cod 457260,
ROMANIA

Hanna Instruments, Inc.

Highland Industrial Park, 584 Park East Drive,
Woonsocket, RI 02895, USA
Phone: +1 401 7657500



Test Laboratory

Hanna Instruments SRL

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Test performed by: Jozsa Ervin / ETO Eng. /

[Name / Title of Signatory / Signature]

Declaration authorized by: Vasile Prodan / ETO Responsible /

[Name / Title of Signatory / Signature]

Issue place / Date

Nușfalău / February 05, 2018

[Place / Date in format: Month, dd, yyyy]