

Document reference no.: 172015/1

Report no.: RN-0355



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

BL983313

Conductivity Mini Controller (0 - 1999 µS/cm), 12Vdc or 115V/230V

Are in compliance with:

Directive 2011/65/EU

Directive 2014/30/EU

Directive 2014/35/EU

RoHS2

EMC and its standards Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or reference s to the other technical specification:

EN 50581:2012

Restriction of the use of certain hazardous substances (RoHS)

EC Regulation 1907/2006 EN 61326-1: 2013 Registration, Evaluation, Authorization and Restriction of Chemicals (REACH)
Electrical Equipment of Measurement, Control and Laboratory use

- EMC requirements Part 1: General requirements and its standards:

EMC requirements Part 1: EN 61000-4-2: 2009, EMC

EN 61000-4-3: 2006, EMC

General requirements and its standards: Electrostatic Discharge Immunity Test

Radiated RF, Electromagnetic Field Immunity Test

4 kV / 8 kV contact / air 10 V / m (80 Mhz to 1 GHz)

Class A

EN 61010-1: 2010 (IEC 61010-1: 2010)

Safety Requirements for Electrical Equipment for Measurement, Control and

Laboratory Use Part 1: General Requirements.

EN 61000-4-4: 2012, EMC

Electrical Fast Transient/Burst Immunity Test

1 kV (5/50 ns, 5 kHz)

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

S.C. Hanna Instruments S.R.L.

Str. Hanna Nr.1, loc. Nușfalău, jud. Sălaj, cod 457260,

ROMANIA

Hanna Instruments, Inc.

Highland Industrial Park, 584 Park East Drive,

Woonsocket, RI 02895, USA

Phone: +1 401 7657500

CE

Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Hanna Instruments SRL

Test performed by:

Jozsa Ervin / ETO Eng. /

[Name / Title of Signatory / Signature]

Declaration authorized by: Vasile Prodan / ETO Responsible /

[Name / Title of Signatory / Signature]

Issue place / Date

Nușfalău / February 02, 2018

[Place / Date in format: Month, dd, yyyy]

FO-CETEST r.3 - 2018.01.30

Page 1 Of 1