

Document reference no.: 152624/2

Report no.: RN-0296



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

BL983321

TDS Mini Controller (0.00 to 19.99 ppm), 12Vdc or 115V/ 230V

Are in compliance with:

Directive 2011/65/EU

Directive 2014/30/EU

Directive 2014/35/EU

RoHS2

EMC and its standards

Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or reference s to the other technical specification:

EN 50581:2012

Restriction of the use of certain hazardous substances (RoHS)

EC Regulation 1907/2006

Registration, Evaluation, Authorization and Restriction of Chemicals (REACH)

EN 61326-1: 2013

Electrical Equipment of Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements and its standards:

General requirements and its standards:

EMC requirements Part 1: EN 61000-4-2: 2009, EMC

Electrostatic Discharge Immunity Test

4 kV / 8 kV contact / air

Class A

EN 61000-4-3: 2006, EMC

Radiated RF, Electromagnetic Field Immunity Test

10 V / m (80 Mhz to 1 GHz)

EN 61010-1: 2010 (IEC 61010-1: 2010)

Safety Requirements for Electrical Equipment for Measurement, Control and

Laboratory Use Part 1: General Requirements.

EN 61000-4-4: 2012, EMC

Electrical Fast Transient/Burst Immunity Test

1 kV (5/50 ns, 5 kHz)

 $The {\it manufacturer} is not {\it liable for modifications made by the user which may alter the device's conformity with the {\it CE label}, nor the consequences thereof.}$

This declaration is the responsibility of the manufacturer:

S.C. Hanna Instruments S.R.L.

Str. Hanna Nr.1, loc. Nușfalău, jud. Sălaj, cod 457260,

ROMANIA

Hanna Instruments, Inc.

Highland Industrial Park, 584 Park East Drive,

Woonsocket, RI 02895, USA

Phone: +1 401 7657500



Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Hanna Instruments SRL

Test performed by:

Jozsa Ervin / ETO Eng. /

[Name / Title of Signatory / Signature]

Declaration authorized by: Vasile Prodan / ETO Responsible /

[Name / Title of Signatory / Signature]

Issue place / Date

Nusfalău / February 02, 2018

[Place / Date in format: Month, dd, yyyy]

FO-CETEST r.3 - 2018.01.30

Page 1 Of 1