

Document reference no.: 170103/1

Report no.: RN-0350



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

HI83300 Multiparameter Benchtop Photometer and pH Meter, 115V or 230V

HI83303 Aquaculture Photometer, 115V or 230V

HI83305Boilers and Cooling Towers Photometer, 115V or 230VHI83306Environmental Analysis Photometer, 115V or 230VHI83308Water Conditioning Photometer, 115V or 230V

HI83314 Wastewater Multiparameter with COD Benchtop Photometer and pH Meter,

115V or 230V

HI83325 Nutrient Analysis Photometer, 115V or 230V **HI83326** Pool and Spa Photometer, 115V or 230V

HI83399 Water and Wastewater Multiparameter Photometer and pH Meter, 115V or 230V

Are in compliance with: Directive 2011/65/EU RoHS2

Directive 2014/30/EU EMC and its standards
Directive 2014/35/EU Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or reference s to the other technical specification:

EN 50581:2012 Restriction of the use of certain hazardous substances (RoHS)

EC Regulation 1907/2006 Registration, Evaluation, Authorization and Restriction of Chemicals (REACH)

EN 61326-1: 2013 Electrical Equipment of Measurement, Control and Laboratory use Class A

EMC requirements Part 1: General requirements and its standards:

EMC requirements Part 1: General requirements and its standards:

EN 61000-4-2: 2009, EMC Electrostatic Discharge Immunity Test 4 kV / 8 kV contact / air EN 61000-4-3: 2006, EMC Radiated RF, Electromagnetic Field Immunity Test 10 V / m (80 Mhz to 1 GHz)

EN 61010-1: 2010 (IEC 61010-1: 2010) Safety Requirements for Electrical Equipment for Measurement, Control and

Laboratory Use Part 1: General Requirements.

EN 61000-4-4: 2012, EMC Electrical Fast Transient/Burst Immunity Test 1 kV (5/50 ns, 5 kHz)

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

S.C. Hanna Instruments S.R.L.

Str. Hanna Nr.1, loc. Nușfalău, jud. Sălaj, cod 457260,

ROMANIA

Hanna Instruments, Inc.

Highland Industrial Park, 584 Park East Drive,

Woonsocket, RI 02895, USA Phone: +1 401 7657500

Test Laboratory Hanna Instruments SRL

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

FO-CETEST r.3 - 2018,01.30

Test performed by: Jozsa Ervin / ETO Eng. /

[Name / Title of Signatory / Signature]

[voller / factor of digitatory / digitatory / digitatory

Declaration authorized by: Vasile Prodan / ETO Responsible /

[Name / Title of Signatory / Signature]

Issue place / Date Nuṣfalău / February 06, 2018

[Place / Date in format: Month, dd, yyyy]

Page 1 Of 1